

**SPECIFICATION SHEET**

<b>Product Name:</b>	CA50 Antigen Calibrator Grade	<b>Catalog #:</b>	L2C009
<b>Description:</b>	Human Cancer Antigen50 Antigen, Calibrator Grade		
<b>Source:</b>	Cell culture supernatant from the human colorectal adenocarcinoma cell line.		
<b>Format:</b>	Calibrator Grade, Liquid		
<b>Concentration:</b>	500KUnits/ml ( CanAg EIA)	<b>Lot #:</b>	L170201
<b>Buffer:</b>	20Mm PBS, pH 7.4 ± 0.02, Tris-HCl.		
<b>Preservative:</b>	0.05% Sodium Azide		
<b>Applications:</b>	Specific methodologies have not been tested using this product.		
<b>Storage:</b>	Short-term store at 2-8°C. Long term store at -20°C		
<b>Inactivation:</b>	Not applicable.		
<b>Warnings:</b>	All cultures have tested negative for HBsAg, HCV, HIV-1 and HIV-2. No test guarantees a product to be non-infectious. Therefore, all material should be considered as potentially infectious.		

This product contains sodium azide, which has been classified as Xn (Harmful) in European Directive 67/548/EEC in the concentration range of 0.1 – 1.0 %. When disposing of this reagent through lead or copper plumbing, flush with copious volumes of water to prevent azide build-up in drains.